

Certificate of Calibration

Manufacturer: **Radian Research, Inc.**
Instrument Model: **RD-33-333 Dytronic Portable Standard**
Serial Number: 300732
Firmware Revision: 07.10.14
Error Specification: .01% worst case

Customer Name: ENERCOM, S.A.
Address: BALBOA, CALLE AKEE #1511-B
APDO.321
BALBOA-ANCON

P.O.Number: E-1973
RMA / Certificate Number: 24623
Calibration Date: 30-Aug-11

Based on the recommended calibration interval, the next calibration is due on: 29-Aug-12



**Quality Management System
ISO 9001 Certified**

Environmental Conditions
Temperature: 23°C +/- 1°C
Humidity: between 35% and 60%

Radian Research's As-Found Test Results showed this Instrument to be:

New In Tolerance Out of Tolerance Inoperative Limited Calibration
For Out of Tolerance conditions, As-Found Data Reports are furnished.

Radian Research Inc. certifies the instrument listed above meets or exceeds all published specifications and was calibrated in compliance with ANSI/NCSL Z540-1 using applicable Radian Research procedures which meet the requirements of ISO 9001:2000. This instrument was calibrated by a Radian Research RS-703A Syntron Automated Calibration System which is traceable to the National Institute of Standards and Technology (NIST). The RS-703A Calibration System is traceable within the limitations of NIST's services, by accuracies derived from accepted values of natural physical constants, or by accuracies derived from accepted ratio type calibration techniques. The RS-703A Calibration System is cross checked and calibrated on a schedule which is adjusted to maintain required accuracies and traceability.

Software used for Calibration: RS-703A Control Program Rel.04.20.02 May 30, 2006
RS-703A serial number: 703143, 703144, 703146

Applicable Traceability & Report Numbers for References used by Radian's Metrology Lab:

Watt-hour, VA-hour, VAR-hour, Q-hour, Amp-hour,

Volt-hour, Volt-Squared hour, AC Volt

Radian Dytronic Transfer Standards consisting of (3) RD-22-RTS,
Serial Numbers: 200717, 200718, 200719
NIST Test Report Number: 697/280054-10; Calibration Due Date 6-Jan-2012.

Time Base (Frequency)

Arbiter Systems Model 1083B Satellite-Controlled Frequency Standard s/n B1057. GPS controlled system with an uncertainty of 0.000002ppm. No calibration required.

DC Volts

Fluke Model 732B DC Volt Standard s/n 7703004 with an uncertainty of $\pm .1$ ppm.
Fluke Certificate Number 9D5460; Calibration Due Date: 14-July-2012.

Resistance

Guildline Standard Resistor Model 9330/10K s/n 62623, 62624. Guildline Test Certificate Numbers C13274 and C13275; with an Expanded Uncertainty of $\pm .390$ ppm. Calibration Due Date: 25-May-2012.

Other

Hewlett Packard 8 Digit Multi-Meter Model 3458A s/n 2823A02816. Agilent Technologies
Test Certificate Number 53298; Calibration Due Date 2-Mar-2012.

Metrology Laboratory Technician Signature

Scott Blackwell LAB
47

*This report shall not be reproduced, except
in full, without prior written approval of the
Calibration Facility*